

Instruction for use - Oedema gloves

Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the glove. Always check material content below. Contact your prescriber or Catell if questions about the product occur.

Size

MODEL	COMPRESSION CLASS	ART.NO
Catell Ödemhandske ¾ Finger	1-2 (15-25 mmHg)	5050
Catell Ödemhandske Fullfinger	1-2 (15-25 mmHg)	6050
EDEMA Light ¾ Finger	1-2 (15-25 mmHg)	7050, 7051
EDEMA Light Fullfinger	1-2 (15-25 mmHg)	7150, 7151
EDEMA Light Long Fullfinger	1-2 (15-25 mmHg)	7160
EDEMA More ¾ Finger	2 (23-32 mmHg)	8050, 8051
EDEMA More Fullfinger	2 (23-32 mmHg)	8150, 8151

MODEL	SIZE	* DIMENSIONS
Only for EDEMA	XXS	13-15 cm
Catell Ödemhandske, EDEMA	XS	15-18 cm
All models	S	18-20 cm
All models	M	20-22 cm
All models	L	22-24 cm
Only for EDEMA	XL	24-27 cm

*Measure MCP-joints circumference. Measurements should be seen as guidelines.

Note – EDEMA is bilateral. Catell Ödemhandske is available in Left/Right.

Properties

- Seams on the outside to minimize pressure on sensitive points.

Use glove

- In purpose to reduce swelling, as recommended by prescriber.
- With the seams outwards.

Contact your prescriber if

- The swelling increases.
- The glove causes numbness.
- The glove causes pain.

Material

Nylon, lycra. Free from latex.

Washing instructions

Follow the washing instructions below. Use a washing pouch. Let dry horizontally.



Single patient use.

Recycling

Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation 2017:745, product class 1